

MC-01-013 | MIRUS Draining Kit

Step-by-Step instruction MIRUS Controller Draining | Schritt-für-Schritt Anleitung
MIRUS Controller Entleerung | Instructions détaillées étape par étape: évacuer
MIRUS Controller | Passo-passo le istruzioni Lo svuotamento MIRUS Controller



GB

Caution: Device must cool
down completely before
procedure can be performed.
Minimum 45 minutes.

DE


Warnung: Vor Entleerung
muss das Gerät vollständig
abkühlen. Mind. 45 Minuten.

FR

Attention: laisser refroidir
complètement avant de
vider le dispositif.
45 minutes du moins.

IT

Attenzione: Prima di
scaricare deve raffreddare
completamente.
Come minimo 45 minuti.



Manufacturer



Technologie Institut Medizin
GmbH (TIM)
Erfurter Straße 9
D-56626 Andernach
Germany
T.: +49 (0) 2632 98 76 297
F.: +49 (0) 2632 98 76 299
info@tim-gmbh.de
<http://www.tim-gmbh.de>

MC-01-013 | MIRUS Draining Kit

Art.No. MC-01-013 | MIRUS Draining Kit Step-by-Step D.00

